


Certificate of Analysis

SD-SS3E-02-2X-WG Lot # 112717ES



SD-SS3E-02-2X-WG Retention Time Standard for ASTM D2887

Lot #: 112717ES Exp: Nov. 2019

Column: SD-002-EX2M (10m x 0.53mm x 0.88 μ m) **Inlet:** 80°C-350°C at 15°C/min **Carrier:** UHP He; 2 mL/min (0.2 min hold) to 20 mL/min at 50 mL/min²
Oven: 30°C-350°C at 15°C/min (5 min final hold) **Detector:** FID at 360°C **Injection Volume:** 0.2 μ L

 **Separation Systems, Inc.**
Experts in Gas Chromatography

Phone 850-932-1433

Sales@SeparationSystems.com

www.SeparationSystems.com